



ANNEX 6

CODE **GT1571**
OVERALL DIMENSION ... 850x850x1300 mm
WEIGHT 50 kg

IMPACT TEST CSTB

STANDARDS

CSTB CAHIER 3778-2017, ANNEX 6

DESCRIPTION

Instrument for determine the energy of ropture by impact

TECHNICAL SPECIFICATIONS

- Steel structure epoxy powder painted

EQUIPMENT

ANNEX 6:

Instrument used to dermine the heavy shock resistance of unglazed ceramic tiles.
Instrument is composed by:

- Sand Plan of proper dimension for a cement plate of 70x70 cm with thickness of 6 cm
- A vertical rod where a you can regulate a support with magnet with a steel ball run on it
- A steel ball of 510 gr
- Steel chrome ball 510 g