



ANNEX 7

CODEGT1572
OVERALL DIMENSION.....500x550x1800 mm
WEIGHT.....50 kg

IMPACT TEST CSTB WEAK SHOCK

STANDARDS

CSTB CAHIER, 3778-2017, ANNEX 7

DESCRIPTION

Instrument used to determine the weak shock resistance of glazed ceramic tiles according to CSTB CAHIER 3778-2017, Annex 7.

Instrument is composed by:

- Rigid support to sustain a cement sample of 40x40 cm
- A vertical rod where you can regulate a support with a magnet with a steel ball run on it
- A steel ball of 50 gr
- A cement sample plate of 40x40 cm with thickness of 4 cm

TECHNICAL SPECIFICATIONS

- Steel structure epoxy powder painted

EQUIPMENT

- Steel frame for positioning the tile
- Steel chrome ball 50 gr

ACCESSORIES AND SPARE PARTS

GT2098	Kit for Doing CSTB Cahier 3778,2017, ANNEX 6
--------	--