



CODE .....GT1016  
MEASURING RANGE .....0 +199 gloss  
OVERALL DIMENSION .....140x75x34 mm  
88x30x45 mm  
WEIGHT .....700 gr

## GLOSSMETER GM-20-60°

## STANDARD

UNI EN ISO 2813, ASTM D 523, ASTM D 2457, DIN 67530

### DESCRIPTION

Instrument to establishing the specular gloss on the ceramic tiles, building material and finish, quality control of paint or ink, plastics, wood furniture, paper (coloured, rolled, painted), nylon, rubber, leather, photographic films, etc. The 60° setting is used for ceramic tiles, photographic films, marble, stone, plastic, paint, etc. As far as particularly reflecting materials – such as polished stainless steel – the measurement may not be accurate. To obtain an accurate report, it is necessary that the sample test may only be flat surface; irregular or particularly curved surfaces may lead to measuring errors.

## TECHNICAL SPECIFICATIONS

- Optical system: incident angle 20° e 60°
- Measuring range: 0 + 100 gloss (resolution 1)
- Display range: 0 + 199 gloss
- Display: LCD 3 digit, 12 mm high
- Light source: LED (wavelength 890 nm)
- Light receiving part: SPD (silicone photodiode)
- Reading angle: 20° e 60° selectioning
- Reading area: 6 x 3 mm oval
- Reproducibility:  $\pm 5\%$  f.s. + 1 digit
- 6 keys function: On, Off, 20°/60° Selection, Zero Cal., Span Cal., Hold
- Power Source: A3 dry-cell battery x 4

## EQUIPMENT

- Calibration plate 20°, 60° and 85°
- Probe
- Connection cable
- Case