

CLIMATIC CELL FROST RESISTANCE - 800

STANDARD

UNI EN ISO 10545 -12 ,EN 202, ASTM C-1026,GOST 27180

DESCRIPTION

Instrument for frost test resistance on ceramic tiles, roof or a similar products. The samples are tested with a programmable series of thermal cycles (- $5 \div +5$ °C and -15 $\div +15$ °C) by repeated immersions or water spraying (the cycles are stored into the microprocessor).

The climatic chamber is programmed to do the UNI EN ISO 10545-12 and EN202 Standards. Other standard on request.

TECHNICAL SPECIFICATIONS

- Steel carrying framework epoxy powder painted, thermal insulated by high density boards
- · Stainless steel sliding door thermal insulated to access chamber test
- Stainless steel chamber test equipped with a forced ventilation system for the uniform temperature distribution onto the samples
- Porthole Ø 200 mm and lighting for inside inspection during the test
- Stainless steel sample basket
- Thermoregulator and probe Pt100 inserted into a ceramic tile for temperature control during the test
- Stainless steel tank for water circulation during the test equipped with heating elements and probe Pt100
- · Hydraulic circuit with integrated pump to spray the water on specimens
- · Microprocessor programmer for automatic execution of the he running thermal cycle
- Pre-setting ISO 10545-12 and EN-202 cycles
- · Chart-recorder fot the control of the running thermal cycle.
- · Supply: 400 V- 50 Hz three phase + neutral (other voltage on request)
- "Touch screen control panel with TFT color to 3,5"
- Saving data and parameters on USB key

ACCESSORIES AND SPARE PARTS

GT0126	Probe Pt 100
GT1323	Paper roll for chart-recorder

٩

CLIMATIC CELL FROST RESISTANCE



12 Gabbrielli Technology s.r.l. I. - Via delle Bartoline 43 - 50041 - Calenzano - (Fi) - Italy - T: +39 055 8811113 - F: +39 055 8861551 - www.gabbrielli.com-info@gabbrielli.com - Rev:0 Copyright © Gabbrielli Technology S.r.l.